



California Medical Device Recall Information



Recall Name

Vycor Medical Recalls the Vycor Viewsite Brain Access System (VBAS) Due to Unidentified Contamination

Recall Date	Product Description	Recalling Firm	Recall Reason
8/21/12	Vycor Viewsite Brain Access System (VBAS)	Vycor Medical, Inc. Boca Raton, FL	<i>Suspected of contamination with an unidentified black fiber</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Vycor Viewsite Brain Access System (VBAS) Model # TC171105 Suspected Lot Recalled: <ul style="list-style-type: none">• VM83450	CA , nationwide	Distributed from June 8, 2012 through July 9, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm337363.htm>